



# FocESS™ Sinuscope Sterilization Tray

3484-001

## INSTRUCTIONS FOR USE

The instructions provided within have been validated by the device manufacturer as being capable of reprocessing reusable medical devices.

Individual sterilizers, instrument cleanliness, specific loading of instrument trays, types and geometry of instruments, sterilization containers, filters and wrappings vary at each location.

### TABLE OF CONTENTS

SUBJECT	PAGE
<b>Read First</b>	
Indications for Use	1
Device Description	1
Sterilization Methods and Configurations	1
Limitations	1-2
Warnings	2
Universal Precautions	2
Points of Use	2
Cleaning	2
Disinfection	2
Maintenance, Inspection and Testing	2
Sterilization	2
Storage	3
Warranty	3

### READ THIS SECTION BEFORE PLACING PRODUCT INTO SERVICE

#### Indications for Use:

The FocESS™ Sinuscope Sterilization Tray is used to organize and protect the FocESS Sinuscopes that are sterilized by a healthcare provider. The FocESS Sinuscope Sterilization Tray is intended to allow sterilization of the enclosed medical devices during these sterilization cycles :

- pre-vacuum steam
- ethylene oxide
- STERRAD® 100S Standard
- STERRAD® NX Standard
- STERRAD® 100NX Standard
- STERRAD® 100NX Express

The Tray is intended to be used in conjunction with a legally marketed wrap. The Tray is not intended on its own to maintain sterility.

#### Device Description:

The FocESS Sinuscope Sterilization Tray is used to enclose and hold the FocESS Sinuscopes in an organized manner during the sterilization process and subsequent storage and transportation. The tray does not have direct patient contact. The tray by itself does not maintain sterility. The tray is composed of aluminum and has a rectangular base with latchable cover. The tray has perforations to allow sterilant penetration. The tray contains silicone instrument holders in the base and/or cover to hold, organize and protect the surgical instruments within the tray during the sterilization process and subsequent storage and transportation.

#### Sterilization Methods and Configurations

##### » Autoclave Sterilization Parameter:

Cycle : Pre-vacuum  
 Temperature : 270° F (132° C)  
 Exposure Time : 4 minutes  
 Minimum Dry Time : 30 minutes

##### » Ethylene Oxide (EO):

Preconditioning Parameters:  
 Temperature: 131°F (55°C)  
 Relative Humidity: 70± 15%  
 Time: 1 hour

##### Sterilization Parameter:

Temperature: 131°F (55°C)  
 Minimum Exposure Time: 120 minutes  
 Minimum Aeration Time: 12 hours

- » STERRAD® 100S Standard Cycle
- » STERRAD® 100NX Standard Cycle
- » STERRAD® NX Standard Cycle
- » STERRAD® 100NX Express Cycle

#### Limitation on Processing

- 1) The end of useful life on the Tray is a minimum of 25 sterilization cycles. Inspect the tray before use for wear and damage caused by use and discontinue use if visible signs of wear are present including corrosion, mechanical failures, cracking, peeling, flaking, broken welds, damaged feet, damaged latches, damaged Hold-It / Hold-Down, discoloration, etc.
- 2) See **Table 2** for sterilization methods and configurations.
- 3) DO NOT OVERLOAD the Tray or components.
- 4) DO NOT OVERLOAD individual Hold-It slots. Load only one instrument per Hold-It slot.
- 5) Inside of sterilizers, DO NOT STACK individually wrapped Trays or components. Separate wrapped Trays or components from each other or any other items on separate shelves of the sterilizer to allow for maximum sterilant flow.

### Limitation on Processing (CONT.)

- 6) The use of non absorbent tray liners (i.e. silicone fingered organizing mat) can cause condensate to pool. If visible moisture is present, re-sterilize with a longer dry time.
- 7) No lumen claims as the FocESS Sinuscope has no lumens.
- 8) The total weight of the system (e.g. tray and instrument load) must not exceed 25 pounds.
- 9) The FocESS Sinuscopes should be prepared and sterilized according to the instructions for use, see Instructions for Use, (REF 3260-001).

### WARNINGS

Use only neutral (6.0 – 8.5) pH detergents to avoid damaging the finish. A detergent with a highly acidic or highly alkaline pH could permanently damage the anodized aluminum finish of the tray and metal components.

**Note : Clean and inspect tray according to the instructions provided prior to placing into service.**

### UNIVERSAL PRECAUTIONS

- Personnel should wear all personal protective clothing and equipment as required by their employers / department's operating procedures for the contamination level they will be exposed to
- Keep dissimilar metals separated during sterilization to prevent corrosion

### POINT OF USE

Remove soil with disposable cloth/paper wipe.

### CLEANING

Equipment: Enzymatic detergent, brush and running water. Use only neutral (6.0 – 8.5) pH solutions, mildly alkaline and free of sodium carbonate to avoid damaging finish for aluminum cassettes. DO NOT use scouring pads or abrasive cleaners. DO NOT store cassette in liquid.

#### Manual Gross Decontamination:

- 1) Remove all visible soil and contaminants using a soft bristle brush. The entire cassette should be immersed while cleaning to aid in the removal of contaminants and to reduce splashing of detergent on personnel for a minimum of 2 minutes.
- 2) Rinse thoroughly for a minimum of 1 minute with clean water to remove all detergent. See rinsing instruction on the detergent label.

#### Ultrasonic clean:

Prepare enzyme wash in an ultrasonic cleaning unit. Place a single cassette in the detergent and run for a minimum of ten minutes. Rinse for a minimum of 2 minutes with cold tap water. Visually inspect cassette for contaminants. Repeat the cycle if necessary to remove visible contamination.

#### Automated washer:

The Trays will complete the cleaning process in an automatic wash system. The validated cleaning process is reflected in **Table 1**.

**Table 1 :**

CYCLE	WATER TEMP.	CLEANING PROCESS
Pre-wash 1 & 2	Cold Tap Water	Re-circulation Time : 2 minutes
Enzyme Wash :	Hot Tap Water	Soaking Time : 4 minutes
Wash 1 :	Heated at 65.5°C	Re-circulation Time : 15 minutes
Rinse 1 & 2	Hot Tap Water	Re-circulation Time : 5 minutes

**Note:** After completion of an automatic wash cycle, visually inspect the Tray for any remaining visible soil. All visible soil must be removed by hand cleaning, brushing, ultrasonic and additional automatic cycles prior to sending to sterilization.

### DISINFECTION

The tray is intended to be terminally sterilized.

### MAINTENANCE, INSPECTION AND TESTING

The tray may be reused until unacceptable deterioration such as corrosion, cracking, rust, peeling, flaking, discoloration, or mechanical failure occurs. Signs of mechanical failure include:

- Broken or cracked corners or welds
- Missing, torn or cut silicone inserts
- Broken or non-working latches
- Missing or damaged feet

### STERILIZATION

See **Table 2** for sterilization methods and cycles that have been validated for the sterilization of the FocESS Sinuscope Sterilization Tray. These are only valid with sterilization equipment that is properly maintained and calibrated – AAMI TIR 12 p.37. Minimum dry times are required to prevent wet packs.

## STERILIZATION (CONT.)

Table 2 :

STERILIZATION METHOD	CYCLE (times)
Steam - PreVacuum	<u>Parameter :</u> Temperature 270° F (132° C) Expose Time 4 minutes Dry Time 30 minutes
	<u>Preconditioning :</u> Temperature 131°F (55°C) Relative humidity 70±15% Precondition time 1 hour
	<u>Sterilization :</u> Exposure time 120 minutes Temperature 131°F (55°C) Aeration time 12 hours
STERRAD 100S	Standard
STERRAD 100NX	Standard
STERRAD 100NX	Express
STERRAD NX	Standard

The following sterilization methods have been validated for the FocESS Sinuscope Sterilization Tray:

- The 4 minute autoclave sterilization cycle in legally marketed wrap, cleared by the FDA.
- The Ethylene Oxide (EO) sterilization cycle in legally marketed wrap cleared by the FDA.
- The STERRAD 100S Standard sterilization cycle in legally marketed wrap cleared by the FDA.
- The STERRAD 100NX Standard sterilization cycle in legally marketed wrap cleared by the FDA.
- The STERRAD 100NX Express sterilization cycle in legally marketed wrap cleared by the FDA.
- The STERRAD NX Standard sterilization cycle in legally marketed wrap cleared by the FDA.

Use an FDA cleared wrap to maintain sterility.

Please consult the sterilizer instruction manual to ensure intended loads are compatible with the intended sterilization cycle.

## STORAGE

Store terminally sterile trays that are wrapped on storage shelf in a horizontal position. Consult wrap manufacturer for shelf life information.

## WARRANTY

### LIMITED WARRANTY FOR THE TRAY.

THIS LIMITED WARRANTY AND THE REMEDY PROVIDED HEREIN ARE EXCLUSIVE AND IN LIEU OF ALL OTHER EXPRESS WARRANTIES AND, UNLESS STATED HERE-IN, ANY STATEMENTS OR REPRESENTATIONS MADE BY ANY OTHER PERSON OR FIRM ARE VOID. THE DURATION OF ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE SHALL BE LIMITED TO THE DURATION OF THE EXPRESS LIMITED WARRANTY. NEITHER SUMMIT MEDICAL, INC. NOR ITS AFFILIATES SHALL BE LIABLE FOR ANY INCIDENTAL, CONSEQUENTIAL OR SPECIAL LOSSES OR DAMAGES, RESULTING FROM THE USE OR INABILITY TO USE THE TRAY, WHETHER RESULTING FROM BREACH OF WARRANTY OR ANY OTHER LEGAL THEORY.

This Limited Warranty gives you specific legal rights, and you may also have other rights which vary from State to State. Some States do not allow limitations on how long an implied warranty lasts, or do not allow the exclusion or limitation of incidental or consequential damages, so the above limitation or exclusions may not apply to you.

**What Is Covered.** Summit Medical warrants the original purchaser that the Tray enclosed with this Limited Warranty conforms to the manufacturer's specifications and is free from defects in workmanship and material for a period of 12 months from the date of original purchase. If the original purchaser transfers the tray to another party, this Limited Warranty will not be enforceable by the party to whom the product is transferred.

**What We Will Do To Correct Problems.** Should your Tray prove defective during this period, you must notify Entellus Medical. You must permit the manufacturer to make such investigation, examination and tests as the manufacturer deems appropriate and, if requested to do so, you will return the product to the address set forth below. Entellus Medical's sole obligation under this Limited Warranty is, at its option, to replace the defective product or products, without charge for parts or labor. Postage, insurance or shipping costs incurred in presenting your tray product for warranty service are your responsibility.

**What Is Not Covered.** This Limited Warranty is contingent upon proper use and maintenance of the product; it does not cover products that have been improperly shipped, or that have been misused, abused, neglected, or improperly maintained, cleaned or stored, or that have been serviced other than by Summit Medical, Inc. or an authorized distributor or dealer of Summit Medical, Inc., or that have been modified without the express approval of Summit Medical, Inc. Failure to follow the directions in the owner's manual may constitute improper use or maintenance of the product and causes this Limited Warranty not to apply. This Warranty does not extend to normal wear or to replacement items.

If you have questions or claims related to this warranty, contact Entellus Medical.



Manufactured by:

Summit Medical  
815 Northwest Parkway, Suite 100  
St. Paul, MN 55121 | USA

Distributed by:

**Entellus Medical Inc.**  
**3600 Holly Lane North, Suite 40**  
**Plymouth, MN 55447 | USA**  
**Phone : 1.866.620.7615**  
**www.entellusmedical.com**

### Symbol Reference Key

Consult instructions for use	Quantity	Catalog Number	Lot Number	Date of Manufacture	Non-Sterile	Caution

Not made with natural rubber latex.